

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 15 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130 178 ✓
Name		James Kimura	ID# 5677 Date 4/28/09 ✓
A	Agency	Alaska State Troopers	Phone # 907 225-5118
Instrument Location		Ketchikan Post	✓
Alco S/N		X172968 ✓	Target Value .083 ✓
			High Pressure 1,100 PSI
B	Alco Test Values	.084 ✓ 1 st Alco	.085 ✓ 2 nd Alco
Signature		James Kimura	4/28/09 ✓
		(OVER)	8mb 5/20/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 178 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

APRIL 28, 2009

OPERATOR'S NAME:

KIMURA/JAMES/K

OPERATOR'S NUMBER: 5677

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: 0000000

DEPT/AGENCY: KETO

CASE/REPORT: 09-000000

TEST TYPE: U

ALCO TARGET VALUE: .083 ✓

ALCO S/N: X172968 ✓

BREATH ANALYSIS

.083 ADJUSTED FOR 30.48 in ✓

ALCO TARGET .084 16:49

BLANK TEST .000 16:50

INTERNAL STANDARD VERIFIED 16:50

ALCO TV 30.46 in .084 16:51

BLANK TEST .000 16:52

SUBJECT SAMPLE .000 16:52

BLANK TEST .000 16:53

ALCO TV 30.48 in .085 16:53

BLANK TEST .000 16:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178

APRIL 28, 2009 ✓

TIME 16:47

DIAGNOSTIC CHECK

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.48 in

FLOW DETECTOR: OKAY

PUMP

WHEEL SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ\]^_`abcdefgijklmnop
qrstuvwxyz{|}~■